



UTTAR PRADESH MEDICAL SUPPLIES CORPORATION LTD.

Regd. Office: SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010

Ph: 0522 2838102 e-mail: drugs@upmsc.in

Result of Technical Evaluation of Tender No. UPMSCL/Drugs-217/09

Date: 28.05.2025

S. No	Bidder Name	Firm Status	Item Code	Item Name	Item technical Status	Sample Status	Final item status based on technical and sample	Remarks for eligible/not eligible
1	AGRINEX CORPORATION	Not Eligible	STK001	Salt Testing Kit (Each Salt Testing kit should have 20ml testing solution or testing capacity of 75-100 samples) [STK001] (NE)	Not Eligible	Sample not submitted.	Not Eligible	1. Cancelled cheque not submitted. 2. In Format XI photo not attached. 3. Format XIII not submitted. 4. Format XVII is not as per tender format. 5. 60 days production capacity not submitted. 6. Average annual turnover is less than minimum requirement, So turnover criteria not fulfilled. 7. ATC UDIN is not explicit, so could not be verified through ICAI. 8. Turnover figures not matching (ATC vs FS) for the F.Y 20-21. 9. Sample not submitted.
2	Alpine Biomedicals Pvt. Ltd	Not Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Not Eligible	Sample pass	Item not eligible	1. In format XI witness sign is not present. 2. ATC issued specifically for the other than this tender. 3. Product permission issued on 2022 hence 3 years market not established.
		Not Eligible	RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-C	Not Eligible	Sample pass	Item not eligible	
3	AVANTOR PERFORMANCE MATERIALS INDIA PVT LTD.	Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Eligible	Sample pass	Eligible	Eligible
		Eligible	RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-C	Eligible	Sample pass	Eligible	Eligible
4	Avience Biomedicals Pvt. Ltd.	Not Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Not Eligible	Sample not submitted.	Item not eligible	1. Non conviction certificate is not valid. 2. Format XI not on Non judicial stamp paper & photo not attached. 3. UDIN is not mentioned on all years F.S. 4. Tender fees & EMD not deposited. 5. Product permission issued on 03.02.2023 hence 3 years market not established. 6. Sample not submitted.
5	Bioline Diagnostics LLP	Not Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Not Eligible	Sample pass	Item not eligible	1. Format XI is not on non judicial stamp paper. 2. Details of turnover for the financial year 20-21 & 21-22 not mentioned on the ATC UDIN verified through ICAI. 3. UDIN not mentioned on Fin. Statement for F.Y 20-21. 4. ATC vs Fin statement T/o figure for the F.Y 22-23 is not
		Not Eligible	RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-C	Not Eligible	Sample pass	Item not eligible	

S. No	Bidder Name	Firm Status	Item Code	Item Name	Item technical Status	Sample Status	Final item status based on technical and sample	Remarks for eligible/not eligible
								<p>matching.</p> <p>5. Documents in support of point 12 of technical specification not submitted.</p> <p>6. Market standing certificate submitted for one year only.</p> <p>7. Test report in support of sensitivity & specificity not submitted.</p> <p>8. Documents as per point 4 of technical specification not submitted.</p>
6	Gene Bio Healthcare Pvt. Ltd	Not Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Not Eligible	Sample pass	Item not eligible	<p>1. Production capacity certificate is not submitted.</p> <p>2. Photo, sign, company name & company seal not done on format VII.</p> <p>3. Format XI is not on non judicial stamp paper & photo not attached.</p> <p>4. In format XII (BOQ) rate of item is mentioned in technical bid.</p> <p>5. Financial statement with UDIN is not submitted.</p> <p>6. Market standing certificate submitted for one year only.</p>
		Not Eligible	RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-Cs	Not Eligible	Sample fail (Samples showed concordant results for negative samples but tend to miss out a fair number of positive samples, showing false negative results)	Item not eligible	
7	GLOBAL BIOSCIENCE S	Not Eligible	STK001	Salt Testing Kit (Each Salt Testing kit should have 20ml testing solution or testing capacity of 75-100 samples) [STK001] (NE)	Not Eligible	Sample pass	Item not eligible	1. Financial statements are not submitted for all F.Y. hence turnover criteria not fulfilled.
8	Medsorce Ozone Biomedicals Pvt. Ltd	Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Eligible	Sample pass	Eligible	Eligible
		Eligible	RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-C	Eligible	Sample fail (Samples showed concordant results for negative samples but tend to miss out a fair number of positive samples, showing false negative results)	Not Eligible	Sample fail (Samples showed concordant results for negative samples but tend to miss out a fair number of positive samples, showing false negative results)
9	Meril Diagnostics Pvt Ltd	Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Eligible	Sample pass	Eligible	Eligible
			RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-C	Eligible	Sample pass	Eligible	Eligible
10	M/s Biogenix Inc.	Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Eligible	Sample pass	Eligible	Eligible
			RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-C	Eligible	Sample fail (Samples showed concordant results for negative samples but tend to miss out a fair number of positive samples, showing false negative results)	Not Eligible	Sample fail (Samples showed concordant results for negative samples but tend to miss out a fair number of positive samples, showing false negative results)

S. No	Bidder Name	Firm Status	Item Code	Item Name	Item technical Status	Sample Status	Final item status based on technical and sample	Remarks for eligible/not eligible
						samples but tend to miss out a fair number of positive samples, showing false negative results)		
11	Oscar Medicare Pvt. Ltd.	Not Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Not Eligible	Sample fail (Samples showed concordant results for negative samples but tend to miss out a fair number of positive samples, showing false negative results)	Not Eligible	1. Format XVII is not as per tender format. 2. Cancelled cheque is not submitted. 3. In format XI photo not attached. 4. UDIN of ATC is wrongly mentioned so the same could not be verified through ICAI. 5. Sample fail (Samples showed concordant results for negative samples but tend to miss out a fair number of positive samples, showing false negative results)
		Not Eligible	RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-C	Not Eligible	Sample fail (Samples showed concordant results for negative samples but tend to miss out a fair number of positive samples, showing false negative results)	Not Eligible	
12	PLASTI SURGE INDUSTRIES PVT LTD	Eligible	STK001	Salt Testing Kit (Each Salt Testing kit should have 20ml testing solution or testing capacity of 75-100 samples) [STK001] (NE)	Eligible	Sample pass	Eligible	Eligible
13	Q-Line Biotech Private Limited	Not Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Not Eligible	Sample pass	Not Eligible	1. Non conviction certificate, product permission market standing certificate & 60 days production capacity is not submitted. 2. In format XI photo not attached. 3. Fin. Statement (Notes to account) for the F.Y 2022-23 not readable. 4. T/o figures for the F.Y 2020-21 is not matching with Fin. statement.
		Not Eligible	RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-C	Not Eligible	Sample pass	Not Eligible	
14	SD Biosensor Healthcare Pvt. Ltd.	Eligible	RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-C	Eligible	Sample fail (Samples showed concordant results for negative samples but tend to miss out a fair number of positive samples, showing false negative results)	Not Eligible	Sample fail (Samples showed concordant results for negative samples but tend to miss out a fair number of positive samples, showing false negative results)

S. No	Bidder Name	Firm Status	Item Code	Item Name	Item technical Status	Sample Status	Final item status based on technical and sample	Remarks for eligible/not eligible
15	Titan Biotech Limited	Eligible	STK001	Salt Testing Kit (Each Salt Testing kit should have 20ml testing solution or testing capacity of 75-100 samples) [STK001] (NE)	Eligible	Sample pass	Eligible	Eligible
16	Transasia Bio Medicals Ltd	Eligible	RDT (HEPA -B)	Rapid Diagnostic Kit For Hepatitis-B	Eligible	Sample pass	Eligible	Eligible
		Eligible	RDT (HEPA -C)	Rapid Diagnostic Kit for Hepatitis-C	Eligible	Sample pass	Eligible	Eligible
17	VISHAL PHARMACEUTICAL LABORATORIES	Eligible	STK001	Salt Testing Kit (Each Salt Testing kit should have 20ml testing solution or testing capacity of 75-100 samples) [STK001] (NE)	Eligible	Sample pass	Eligible	Eligible

Managing Director
UPMSCL