



Uttar Pradesh Medical Supplies Corporation Limited

(A Govt. of Uttar Pradesh Undertaking)

(CIN: U85310UP2018SGC102425)

Registered Office : SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226010

E-Mail: equipment@upmsc.in Website: www.upmsc.in Contact Number: 0522-2838102

Date- 20.01.2026

Technical Corrigendum:-1

Tender Reference No.: GEM/2025/B/6950172

Name of Equipment: Brain Stem Evoked Response Audiometer with ASSR with OAE

Sr. No.	Reference of Tender Clause	Tender Clause	Revised Clause
1.	Technical (BERA) Specification	Input Voltage range should be 20 μ V to 50mV	Deleted
2.	Technical (BERA) Specification	Sampling rate should be 200Hz- 30000 Hz	Sampling rate should be 200Hz- 30000 Hz or better
3.	Technical (BERA) Specification	A/D converter should be 16 bits	A/D converter should be 16 bits or better
4.	Technical (BERA) Specification	Impedance measurement range should be 0–25k Ω	Impedance measurement range should be 0.5–25k Ω
5.	Technical (BERA) Specification	Notch Filter type should be Adaptive and Recursive.	Deleted
6.	Technical (BERA) Specification	Notch filter 50 or 60Hz (switchable) should not be less than 40dB.	Deleted
7.	Technical (ASSR) Specification	Range of stimulation level (for click) should be 0–130 dB SPL for Insert Ear Phone	Range of stimulation level (for click) should be 0–100 dB NHL for Insert Ear Phone
8.	Technical (ASSR) Specification	Range of stimulation level (for tone burst) should be 0–130 dB SPL for Insert Ear Phone	Range of stimulation level (for tone burst) should be 0–100 dB NHL for Insert Ear Phone
9.	Technical (ASSR) Specification	Stimulus repetition rate should be 0.1–90Hz with the admissible relative deviation of $\pm 0.1\%$.	Stimulus repetition rate should be 10–100Hz with the admissible relative deviation of $\pm 1\%$.
10.	Technical (ASSR) Specification	Tone burst stimulation tone range should be 50–10000Hz.	Tone burst stimulation tone range should be 250–4000Hz or more
11.	Technical (ASSR) Specification	Stimulus Duration of Tone burst stimulus should be in Millisecond	Deleted
12.	Technical (ASSR) Specification	Frequency Range 250Hz - 5000Hz and above	Frequency Range 250Hz - 4000Hz or above



Uttar Pradesh Medical Supplies Corporation Limited

(A Govt. of Uttar Pradesh Undertaking)

(CIN: U85310UP2018SGC102425)

Registered Office : SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226010

E-Mail: equipment@upmsc.in Website: www.upmsc.in Contact Number: 0522-2838102

Sr. No.	Reference of Tender Clause	Tender Clause	Revised Clause
13.	Technical Specification (ASSR)	Stimulus Duration of click stimulus should be in Microsecond.	Deleted
14.	Software	Should be able to create customizable and user defined report templates.	Deleted
15.	Software	Should have the feature of Auto marking.	Should have the feature of Auto/Manual marking.
16.	Hardware	Should be Portable and to be supplied with table mounting stand	Should be Portable and to be supplied with table mounting/wall mounting/table top.
17.	Hardware	Unit should have in built Amplifier.	Unit should have in built/external Amplifier.
18.	Hardware	It should take power through USB from Laptop or Desktop	It should take power through USB from Laptop or Desktop or AC Mains
19.	General Specification:-	Must Have Auto Calibration	Deleted
20.	Product/Service specification	Quality Certificate:- USFDA/European CE Standards	<ol style="list-style-type: none">1. EU CE from notified body/USFDA/BIS Certificate.2. CDSCO Licence3. ISO 13485 certified